



**USP <88> Class Plastic Designation**

Sponsor: Garlock Sealing Technologies  
1666 Division Street  
Newark, NY 14522-9350

Attn: Matt Tones

Test Nos.: T09-1343, 1344, 1345, Amendment I

Test Material: GYLON 3500

This test material meets the requirements for USP 27 – NF 22, General Chapter <88>, for Class VI designation at a temperature extraction of 121°C for 1 hour.

*Jonia Schauboeck*

Study Director

*4-8-10*

Date



Amendment of USP Class VI Designation Certificate  
(Reference SRP-005)

Regarding Test Nos.: T09-1343, 1344, 1345

Amendment # 1

The following correction has been made to the above mentioned document:

Amendment 1: The USP – National Formulary editions (USP 27 – NF 22) and general chapter (<88>) have been added to the USP Class Plastics Designation Certificate originally issued by Tonia Schaubroeck on July 6, 2004.

Reason: The USP edition and general chapter governing USP Class Plastics Designation testing were not referenced on the original certificate. This information was added to the document for clarification purposes.

*Tonia Schaubroeck*

Study Director

*4-8-10*

Date